

InterLab®

POLYSEED®



CERTIFICATE OF ANALYSIS

P.O. Box 130549
Spring, Texas 77393
Phone: (281) 298-9410 Fax: (281) 298-9411

Description: GLUCOSE-GLUTAMIC ACID
Catalog Number: P510
Lot Number: H180-04

Mfg Date: 07/24/2018
Expiration Date: 07/24/2020

ANALYTICAL SECTION

Test	Specification	Test Result
Appearance	clear, colorless solution	Pass Test
APHA 5-Day BOD, Method 5210	198.0 mg/L BOD +/-30.5 mg/L	200.7 mg/L BOD
Mfg. with 150 mg/L D-Glucose (CAS 50-99-7) and 150 mg/L L-Glutamic Acid (CAS 56-86-0)		
Each lot certified by an independent NELAP accredited laboratory		
Balances used during manufacture are traceable through NIST Test Number 822-275872-11		

Submitted By: Matthew V. Petrina, Chemist Supervisor
An ISO9001:2008 certified company. Registration # 0306-01