

InterLab®

POLYSEED®



CERTIFICATE OF ANALYSIS

P.O. Box 130549
Spring, Texas 77393
Phone: (281) 298-9410 Fax: (281) 298-9411

Description: GLUCOSE-GLUTAMIC ACID
Catalog Number: P810
Lot Number: H313-20

Mfg Date: 11/29/2018
Expiration Date: 11/29/2020

ANALYTICAL SECTION

Test	Specification	Test Result
Appearance	clear, colorless solution	Pass Test
APHA 5-Day BOD, Method 5210	198.0 mg/L BOD +/- 30.5 mg/L	211.1 mg/L BOD
Mfg. with 150 mg/L D-Glucose (CAS 50-99-7) and 150 mg/L L-Glutamic Acid (CAS 56-86-0)		
Each lot certified by an independent NELAP accredited laboratory		

Submitted By: Greg Albright, Chemist Supervisor
An ISO9001:2008 certified company. Registration # 0306-01